

**OFFICE OF  
THE STATE HEALTH SOCIETY, MANIPUR**

**Sl. No.:**

**Issued To M/s.** \_\_\_\_\_

**NOT TRANSFERABLE**

**TENDER FOR SUPPLY OF MEDICAL EQUIPMENTS  
NATIONAL COMPETITIVE BIDDING**

---

TENDER REFERENCE : 38/03/NRHM-10(pt-I)Fin  
DATE OF COMMENCEMENT OF : 5<sup>th</sup> September, 2012  
SALE OF TENDER DOCUMENT  
LAST DATE FOR SALE OF : 20<sup>th</sup> September, 2012  
TENDER DOCUMENT  
LAST DATE & TIME OF RECEIPT OF : 13.00 hrs of 3<sup>rd</sup> October, 2012  
TENDER  
TIME AND DATE OF OPENING : 15.00 hrs of 3<sup>rd</sup> October, 2012  
OF TENDER  
PLACE OF OPENING OF TENDER : Office of the SMD, Old Secretariat.  
ADDRESS FOR COMMUNICATION : Office of the State health Society  
B.T.Road,F.W.Department bld. Imphal  
Tender bid can also be submitted to  
the Resident Commissioner, Manipur  
Bhavan , New Delhi  
COST OF THE TENDER DOCUMENT : RS.1000/-

**OFFICE OF  
THE STATE HEALTH SOCIETY, MANIPUR**

**TENDER FOR SUPPLY OF MEDICAL EAQUIPMENTS**

1. Sealed tenders in 2-bid system are invited by State Health Society, Manipur for the supply of Medical Equipments as per enclosed specification and terms & conditions from reputed and experienced manufactures/ authorized distributors of similar items and shall be received till 13.00 hrs on 3<sup>rd</sup> October,2012.

**2. ELIGIBILITY CRITERIA –**

- a. Bidder shall be a manufacturer of medical equipment having requisite manufacturing facility.
- b. The bidder shall have market standing continuously for the past 3 years in supplying similar items with customers' satisfaction.
- c. Authorized dealer / distributors of a manufacturer or Indian Agent of an overseas vendor are also eligible to participate in the tender, provided they furnish the authorization for the items and having the dealership/ distributorship for the subject items at least for three years continuously.
- d. The average annual turn over of the bidder during the last three years shall not be less than Rs 1(one) crore.
- e. Supply shall be effected directly by the bidder and not through any other agency

**GENERAL CONDITIONS**

3. A complete set of bidding documents may be purchased by any interested eligible bidder on submission of written application and upon payment of non – refundable fee of Rs.1000/-in the form of Demand Draft in favour of State health Society, Manipur payable at Imphal.

Bidding document may be obtained during office hours on all working days either in person or by post.

4. All bids must be accompanied by Earnest Money Deposit (EMD) as specified in the bid document and must be furnished with the bid.
5. Bids shall be opened in presence of bidders/ representatives who intend to attend, on the specified date and time.
6. At any time prior to the date of submission of bid, the authority inviting tender may, for any reason, whether at his own initiative or in response to a clarification from a prospective bidder, modify the bidding documents by an amendment. All prospective bidders who have received the bidding document will be notified of the amendment in writing and the amendment shall be binding on them. In order to provide reasonable time to take the amendment into account in preparing the bid, the authority inviting tender, may at his discretion, extend the date and time for submission of bids.
7. The bid and all correspondence and documents shall be in the English language.
8. The tender document is not transferable.
9. The tender document shall be signed by the tenderer in all the pages with official seal.
10. Interested eligible bidders, if so desire, may obtain further information from the office of the authority inviting tender.
11. **TECHNICAL BID "COVER 'A'"**

The bidder shall furnish along with the bid, the following, in a separate cover, hereinafter called **Cover 'A'**

- a) All bids must be accompanied by **EARNEST MONEY DEPOSIT** of an amount for each schedule as shown in the list of items, in the form of Demand Draft/Bankers Cheque in favour of State health Society. EMD

submitted in any other form or bids without EMD shall not be accepted. **It is to be noted that separate EMD shall be submitted for each schedule being quoted.**

The EMD of the successful bidder shall be retained till completion of the tender period but shall not carry any interest. If the successful bidder fails to execute the agreement and/ or fails to deposit the performance security within the specified time, or withdraws his bid within the validity period of the bid, the EMD shall be forfeited.

The EMD of the unsuccessful bidders will be returned within 30 days after the finalization of the tender.

- b) Duly attested photocopies of valid manufacturing license for the products offered.
- c) Declaration on details of manufacturing unit, installed capacity of the item quoted, testing facilities and nearest after sales service facility with details of technical personnel, along with non - conviction certificate/ declaration for the past 3 years. (Refer **Annexure I**)
- d) Details of supplies made during the last 3 years with summary of Purchase Orders and performance certificates issued by clients in the specified format (Refer **Annexure II**). Items supplied to Govt. institutions and to programs sponsored by WHO, UNICEF etc., if any for the last 3 years with copies of Purchase Order and Performance certificates are to be separately highlighted.
- e) Annual Turn Over Statement for the last 3 financial years i.e. 2009 – 2010, 2010 – 2011, 2011 – 2012 along with statement of concurrent commitment in the specified format (Refer **Annexure III**) certified by a Auditor/ Chartered Accountant.

- f) Current & Valid Sales Tax Clearance Certificate/ VAT Certificate as on 31.03.2012.
  - g) Authorization like Power of Attorney or Resolution of the Board for the officer of the company who have signed the tender document and the bid
  - h) Undertaking in the form at **Annexure-IV** confirming acceptance of all terms and conditions of the tender including special conditioned at **clause 12**.
  - i) An undertaking on fraud and corruption as per **Annexure-V**.
  - j) In case of dealer/distributor or Indian agent of an overseas manufacturer, Authorization from the manufacturer for the items quoted in **Annexure - VI**.
  - k) Catalogue, literature and schematic diagrams (wherever applicable) of all the equipment being offered.
  - l) The List of items quoted shall be furnished in **Annexure - VII**. The list shall specifically indicate the make/ model no., manufacturer and brand name (if any) along with technical specifications. But this list **shall not indicate prices of the items.**
- 12. In the technical bid, the bidder shall confirm that, in case he becomes the successful bidder he shall abide by the following stipulations which shall also form a part of his undertaking at Annexure IV. [Ref. 11 (h)].**
- 13. Warranty Maintenance:** The medical equipment supplied shall be covered under comprehensive warranty for a period of 3 years from the date of commissioning and handing over and any maintenance or repair arising during this period of 3 years including replacement of parts if any, shall be carried out by the supplier at his own cost without any cost liability on the purchaser.

- a) ANNUAL MAINTENANCE CONTRACT (AMC): The supplier shall quote the %age rates for the Comprehensive Annual Maintenance of the equipment supplied, after expiry of warranty period of 3 years, for a further period of 5 years, in the price bid. Price comparison shall include this 5 years comprehensive maintenance charges. **Price bids without AMC rates shall not be considered for evaluation.** The supplier shall also quote annual maintenance charges for next 5 years both on comprehensive basis (i.e. inclusive of spares) and for services only excluding spares.
- b) Authorization from Equipment Manufacturer: The successful bidder shall ensure that the manufacturers of the equipment being supplied shall ensure full guarantee/ warranty obligations for a period of 3 years from the date of commissioning of the supplied equipment. They shall also confirm that they shall provide Comprehensive Annual Maintenance Services for at least 5 years after expiry of warranty period of 3 years, in terms of Comprehensive Annual Maintenance Contract to be entered into between the supplier/ Manufacturer and purchaser.
- c) Response Time for Breakdown Calls: Maintenance personnel shall attend to the breakdown within 72 hrs of reporting the fault.
- d) Preventive Maintenance: During the warranty period and AMC, in addition to attending break down calls, there shall be regular quarterly preventive maintenance visits.
- e) Operation & Maintenance Manuals: Two sets of operation & maintenance manuals along with circuit diagram and spare parts list of equipment shall be supplied.
- f) Installation, commissioning & Warranty Service: Supply, installation, commissioning and successful trial run of the equipment & subsequent,

warranty / AMC servicing shall be done at site by the supplier themselves or their authorized agent.

- g) Training: If required, free demonstration, operation and maintenance training of the assigned personnel at site shall be provided.

14. All the above documents mentioned in **Para 11 & 12 along with Agreed Terms and Conditions as per Annexure-VIII** shall be submitted in separate sealed cover superscribed as “COVER ‘A’ “, TENDER FOR SUPPLY OF MEDICAL EQUIPMENTS FOR DHs AND CHCs (TENDER NO. 38/03/NRHM-10 PT-I) DUE AT 13HRS ON 3<sup>rd</sup> OCTOBER, 2012 & ADDRESSED TO STATE MISSION DIRECTOR, STATE HEALTH SOCIETY, MANIPUR.

15. PRICE BID “COVER ‘B’”

- a) Bid shall be typed/ written and correction, if any, in the bid shall invariably be attested with full signature by the bidder with date, failing which the bid shall be considered ineligible for evaluation. Corrections done with correction fluid shall also be duly attested.
- b) **The rate quoted per unit shall be the landed cost at destination, inclusive of packing, forwarding, Excise Duty, Sales Tax/VAT, Freight, Insurance, Installation / Commissioning etc and warranty maintenance for 3 years, showing the break – up of cost.**
- c) The landed price per unit including 3 years warranty plus charges for 5 years comprehensive maintenance contract shall be the criteria for determining the L1 rate.
- d) Each page of the price bid shall be duly signed by the bidder with official seal.
- e) The tenderer shall duly fill up **Annexure-X** for items quoted and their rates along with price break-up details as per **Annexure VIII** and shall

submit in a separate sealed cover superscribed as “ **COVER ‘B’** “ ,  
TENDER FOR SUPPLY OF MEDICAL EQUIPMENTS (TENDER NO.  
38/03/NRHM-10 PT-1) DUE AT 13.00HRS ON 3<sup>rd</sup> OCTOBER, 2012 &  
ADDRESSED TO STATE MISSION DIRECTOR, STATE HEALTH  
SOCIETY, MANIPUR.

- f) The **Cover ‘A’ and Cover ‘B’** shall be separately sealed and both these covers shall be put in another sealed cover superscribed as “TENDER FOR SUPPLY OF MEDICAL EQUIPMENT FOR DHs AND CHCs(TENDER NO. 38/03/NRHM-10 PT-1) DUE AT 13.00 HRS ON 3<sup>rd</sup> OCTOBER, 2012 & ADDRESSED TO STATE MISSION DIRECTOR, STATE HEALTH SOCIETY, MANIPUR.
16. Cover ‘A’ i.e. Technical Bids shall be opened at the office of the purchaser or at any other place earlier intimated, on the date and time specified in presence of the intending bidders or their authorized representatives.
17. Cover ‘B’ i.e. the Price Bids of the bidders who meet the eligibility criteria, evaluated based on the details furnished in Cover ‘A’ shall only be opened and the date and time of opening of Cover ‘B’ shall be intimated to the eligible/ short listed bidders

#### **18. TENDER EVALUATION**

Tenders will be evaluated with reference to various criteria and one of such criteria is that the rate per unit (landed price including 3 years warranty and 5 years comprehensive maintenance charges) for determining the L1 rate (Lowest rate). Conditional discounts shall not be taken into account for price comparison. However same shall be considered in case of placing order if the bidder happens to be L1.



## **19. VALIDITY OF BIDS**

Bids shall remain valid for acceptance for a period of 90 days after opening of Technical bid i.e. Cover 'A'. Bids with shorter validity shall be rejected. Purchaser may solicit bidders' consent to an extension of bid Validity period. A bidder may refuse extension request without forfeiting the bid Security.

## **20. VALIDITY OF TENDER**

**The validity of tender of the successful bidder shall be at least 1 year from the date of finalization of the order and the successful bidder(s) are bound to supply the items at agreed rates during this period. This validity period may be further extended with mutual consent.**

## **21. REASONABILITY OF RATES/ FIRM PRICE**

a) The bidder shall certify that the rates quoted are the lowest ones for any institution in the country.

If the bidder is stockiest / distributor/ dealer, he shall confirm that the price quoted are based on manufacturer's list price with appropriate discount & shall enclose manufacturer's price list or priced quotation in support of his claim.

b) During the period of the contract, if the price of any tendered item is reduced due to any reason including any Law or Act of the Central/State Government, the bidder shall be statutorily bound to intimate the reduced rates immediately to the purchaser and shall charge the reduced rates. The purchaser is empowered to unilaterally effect such reduction as is necessary in rates, in case the bidder fails to notify or fail to agree to such reduction of rates.

- c) Subject to the condition stipulated above, the prices shall remain **firm** for the validity period of tender and on no account any increase in price shall be entertained till completion of the tender period.
- d) No bidder will be allowed at any time on any ground whatsoever, to claim revision of or modification in the rates quoted by him. The representation of the bidder that computation/ typographical or clerical error etc. has been committed in the bid and request for reversion on such plea shall not be entertained after opening of the bid.

**22. STATUTORY TAXES/ DUTIES**

In case of any enhancement of Taxes and/ or duties or levy of fresh Taxes/ duties due to Statutory Act of the Govt., after date of submission of the tenders and during the contractual delivery period, additional or fresh levies so imposed will be allowed to be claimed as extra without any change in the price structure approved under the tender. For this purpose, the supplier shall produce a certificate from the authority concerned certifying that the item supplied falls under particular tariff resulting in additional/ fresh levies for the supplied item.

However, the same shall not be borne by the purchaser in case such levies become applicable after expiry of the contractual delivery period stipulated in the contract.

Further, in case a successful bidder has been enjoying Excise Duty exemption on any criteria like turnover etc. and at a later date, during currency of the contract, even if Excise Duty becomes chargeable on goods manufactured, the same shall be to the supplier's account and shall not be borne by the purchaser.

**23. PERFORMANCE SECURITY DEPOSIT:**

The successful bidder, within 15 days of receipt of Purchase Order, shall be required to submit Performance Security Deposit of 10% of the order value in the form of Performance Bank Guarantee in favour of the purchaser valid for a period of 3 years from the date of completion of order. The Bank Guarantee shall be returned on completion of the Warranty period of the goods supplied.

However, if the supplier fails to execute the order or fails to perform the services as per contract, in addition to other penal actions, the Bank Guarantee shall be encashed & the amount forfeited.

**24. AGREEMENT**

The successful tenderer shall execute an agreement on non-judicial stamp paper of value of Rs.100/- (stamp duty to be paid by the tenderer) as per proforma in **Annexure XI** within 15 days from the date of the intimation from Tender Inviting Authority informing that his tender has been accepted.

**25. NON ASSIGNMENT**

The tenderer shall not, at any time, assign, sub-let or make over the contract or the benefit thereof or any part thereof to any person or persons what so ever.

**26. COMMUNICATION**

All notices or communications relating to or arising out of this agreement or any of the terms thereof shall be considered duly served on or given to the tenderer if delivered to him or left at his premises, places of business or abode.

**27. ANNULMENT OF AWARD, FORFEITURE OF SECURITY DEPOSIT & FRESH AWARD**

Failure of the successful bidder to comply with the requirements of signing of contract and / or submission of performance security within the

time schedule as stipulated above shall constitute sufficient grounds for the annulment of the award and forfeiture of the bid security.

Under such a situation, the proposal may be reviewed for award of the contract on the next lowest evaluated technically qualified bidder or go for a fresh bid depending on the circumstance. In case it is decided to go for the next lowest bidder, negotiation may be considered to bring down their price nearer to the originally evaluated & lowest bidder.

**28. TENTATIVE QUANTITY**

The quantity mentioned is only the tentative requirement and may increase or decrease as per the decision of Tender Inviting Authority. The rates quoted should not vary with the quantum of the order or the destination.

**29. INSPECTION & QUALITY ASSURANCE**

- a) The purchaser and /or his authorized representative(s) have the right to inspect the manufacturing facilities of those companies who have quoted or whose items have been quoted for this tender, before accepting their rates or before awarding the contract, or at any point of time during the continuance of the tender and has also the right to reject the tender or not to reorder based on facts brought out during such inspections.
- b) During the process of manufacture/ fabrication of the ordered items, stage wise as well as random inspections may be carried out by authorized technical personnel to ensure compliance to specification/quality.

However, such inspection shall not absolve the supplier from his responsibility of strictly adhering to the specifications & other conditions spelt out in the tender.

- c) In case of major equipment, the bidder shall submit a stage wise quality assurance plan and the final Factory Acceptance Test plan of the complete unit to the tender inviting authority.

**30. DELIVERY CONDITION**

- i) The supply of items and successful commissioning shall be completed within 60 days from the receipt of the Purchase Order.
- ii) The supply, installation, commissioning of the equipment and trial run have to be done at site by the supplier or his authorized agent. No additional charges for these services shall be paid. The supplier or the Indian Agent shall be responsible for these services for imported items.
- iii) The units as per order shall be handed over to the authorized representative(s) of the purchaser at the specified location and the same shall be duly receipted after installation, commissioning and satisfactory demonstration of full functioning of the unit.

**31. PAYMENT TERMS**

No advance payment shall be made. 90% payment for the supplied items shall be made after receipt of the fully functional items and completion of all codal formalities subject to submission of Bank Guarantee for Performance Security, relevant documents, test certificates, warranty certificates etc.

Balance 10% payment shall be released on completion of the total supply and on successful operation of all the units for at least 3 months. Payment for comprehensive Annual Maintenance Contract after 3 years of warranty period shall be made annually i.e. after 4<sup>rd</sup>, 5<sup>th</sup>, 6<sup>th</sup> years and so on at the agreed rates.

**32. PENALTY FOR DELAY IN DELIVERY**

- a) In case there is delay in delivery beyond the stipulated period as mentioned in delivery clause, there shall be reduction in price @ 0.5% of the value of delayed goods per week of delay or part thereof subject to a maximum of 10% of the total order value.
- b) Once the maximum price reduction is reached, termination of the contract may be considered.
- c) Non-performance of the contract provisions shall make the successful bidder liable to be disqualified to participate in any tender for the next 5 years, in addition to forfeiture of EMD and Bank Guarantee for Performance Security and other penal provisions.

**33. FORCE MAJEURE**

The above condition of delivery period, price reduction & termination etc. are subject to force majeure conditions which are beyond the control of the supplier, do not involve fault or negligence of the supplier and are not anticipated. Such events may include but are not limited to riots, mutinies, war, fire, storm, tempest, flood, epidemics, or other exceptional causes like quarantine restrictions, freight embargoes. On specific request made by the tenderer the time period of supply may be extended by the purchaser at his discretion for such period as may be considered reasonable. However, the condition shall not include scarcity of raw materials, power cut, labour dispute, failure of sub-vendor and increase in cost of raw material.

**34. FRAUD & CORRUPTION:**

The bidders, suppliers & contractors shall observe the highest standard of ethics during bidding and during performance of the contract. For the purposes of this provision, the following acts shall be considered as corrupt and / or fraudulent practices -

- a) "Corrupt Practice" means offering, giving, receiving, or soliciting directly or indirectly, of any thing of value to influence the action of an official in the procurement process or in contract execution.
- b) "Fraudulent Practice" means misrepresentation or omission of facts in order to execution of contract.
- c) "Collusive practice" means a scheme or arrangement between two or more bidders, with or without the knowledge of the purchaser, designed to establish bid prices at artificial, non-competitive level.
- d) "Coercive Practice" means harming or threatening to harm, directly or indirectly, persons or their property to influence their participation in a procurement process or in execution of a contract.

During the process of evaluation of a bid or proposal for award of a contract, if it is detected that a bidder directly or through agent has engaged in corrupt, fraudulent, collusive or coercive practice in competing for the contract in question, then a) the bid shall be rejected and b) declare the firm ineligible for a specific period or indefinitely to participate in a bidding process.

In the bid document itself, an undertaking from the bidders may be obtained in the format at **ANNEXURE- V**.

**35. LOCAL CONDITIONS**

It will be imperative on each bidder to fully acquaint himself of all local conditions and factors that would have any effect on performance of the Contract. The purchaser shall not entertain any request for clarifications from the bidder regarding such local conditions nor shall accept any offer conditional to the local factors. No request for any change of price or extension of time schedule of delivery of goods shall be entertained after purchaser accepts the bid.

**36. WAIVAL / ALTERATION ETC**

Bidders request for waival, alteration etc. in respect of bid document fee, EMD, performance security etc. shall not be entertained and hence no formal reply shall be given for such requests.

The un priced bids shall not be opened of those bidders who have not complied with the provisions of the Bid Document Fee and / or EMD clause of the Tender Document.

**37. ADJUDICATION/REVIEW BOARD**

Any dispute arising out of or during execution of the contract shall be settled with mutual agreement through an Adjudication/Review Board appointed by the appropriate authority of the purchasing organization, having officers belonging to other departments not related to the purchasing department.

**38. SAVING CLAUSE**

No suit, prosecution or any legal proceedings shall lie against Tender Inviting Authority or any person for anything that is done in good faith or intended to be done in pursuance of tender.

**39. LAWS GOVERNING THE CONTRACT & JURISDICTION**

The contract shall be governed by the laws in force in India. In the event of any dispute arising out of the tender such dispute would be subject to The contract shall be governed by the laws in force in India. In the event of any dispute arising out of the tender such dispute would be subject to the jurisdiction of the Civil Court within the city of Imphal only.



## List of Equipments/items with specification, quantity

**Medical Equipments :-**

Sl. No.	Item specification with quantity	Quantity needed
1.	<p>Infant Radiant warmer Advanced infant warmer with phototherapy and Trolley Neotherm intelligent infant warmer 6000 series Advanced oval shaped Trolley with Electronic Height adjustment,360 rotation, bed tilting mechanism</p> <p><b>Standard Radiant Heat Warmer with Fixed Body Cradle and Double Surface Phototherapy</b> Micro Computer temperature Controller</p> <ul style="list-style-type: none"> <li>▪ Provides thermal radiation with highest degree of accuracy</li> <li>▪ User friendly feather touch control</li> <li>▪ Safety alarms along with high temp. cut off</li> </ul> <p>Metallic heater box</p> <ul style="list-style-type: none"> <li>▪ Consists of ceramic Infra Red Heater with higher power output and more durability than any tubular Silicon Quartz glass heaters</li> </ul> <p>Halogen Examination Lamp Provides accurate assessment of infant colour and sufficient illumination during procedures</p> <p><b>Baby Tray</b></p> <ul style="list-style-type: none"> <li>▪ Clear acrylic collapsible side panels to assess an infant from all sides</li> <li>▪ Acrylic Baby Tray with Head Up/Down facility</li> <li>▪ Cushion Mattress</li> </ul> <p>Types of Temp. Controllers</p> <ul style="list-style-type: none"> <li>▪ Air/Pre-warm</li> <li>▪ Skin/Manual</li> </ul> <p><b>Air/Prewarm</b> <b>Prewarm mode</b></p> <ul style="list-style-type: none"> <li>▪ Time programmable</li> <li>▪ Heater output settable</li> <li>▪ Over temp. cut off</li> </ul> <p>Timer/APGAR Timer Auto audio alarm mute Mute time programmable Safety alarms (with status LED)</p> <ul style="list-style-type: none"> <li>▪ Probe failure</li> <li>▪ &gt;1.5 °C of set Temp.</li> <li>▪ &lt;3.0 °C of set temp.</li> <li>▪ &gt;39 °C of air temp.(Actual temp starts blinking)</li> <li>▪ Power failure</li> <li>▪ Safety high temp. cut off at 39°C</li> </ul> <p><b>Skin/Manual</b> <b>Manual mode</b></p>	8

	<ul style="list-style-type: none"> <li>▪ Time programmable</li> <li>▪ Heater output settable</li> <li>▪ Over temp. cut off</li> </ul> <p>Timer/APGAR Timer Auto audio alarm mute Mute time programmable Safety alarms (with status LED)</p> <ul style="list-style-type: none"> <li>▪ Probe failure</li> <li>▪ &gt;1.5°C of set temp.</li> <li>▪ &lt;0.5°C of set temp.</li> <li>▪ <span style="float: right;">&gt;38°C of air</span> temp.(Actual temp starts blinking)</li> <li>▪ &lt;34°C skin temp. (Actual temp starts blinking)</li> <li>▪ Power failure</li> <li>▪ Safety high temp. cut off at 38°C</li> <li>▪ If skin probe is to be used for air control, &lt;34°C limit can be programmed at desired air low temp. alarm</li> </ul> <p><b>Skin/Air/Manual (LCD)</b> <b>Manual mode</b></p> <ul style="list-style-type: none"> <li>▪ Time programmable</li> <li>▪ Heater output settable</li> <li>▪ Over temp. cut off</li> </ul> <p>Timer/APGAR Timer Baby attending alarm Auto audio alarm mute Mute time programmable Safety Alarms Messages on LCD Skin Mode</p> <ul style="list-style-type: none"> <li>▪ Air high 1.5°C</li> <li>▪ Air low 3.0°C</li> <li>▪ Air over range</li> <li>▪ Air probe failure</li> </ul> <p>Safety high temp. cut off at</p> <ul style="list-style-type: none"> <li>▪ Skin temp. &gt;38°C</li> <li>▪ Air temp. &gt;39°C</li> </ul> <p>Heater &amp; Power failure alarm</p> <p><b>Infra Red Ceramic Heater Box Assembly</b> Heating Element</p> <ul style="list-style-type: none"> <li>▪ Hermetically sealed</li> <li>▪ Cannot get oxidised, thus provides long life</li> </ul> <p>Heater grill</p> <ul style="list-style-type: none"> <li>▪ Covers only heater element</li> <li>▪ Ease of reflector cleaning</li> <li>▪ Avoids reflector blackening</li> <li>▪ Efficient heater output</li> </ul>	
--	---	--

	<p>230 volts AC</p> <ul style="list-style-type: none"> <li>▪ Not exposed to patient</li> <li>▪ No sparks, ensures 100% safety</li> </ul> <p><b>Silicone Quartz heater box assembly</b> Heating Element</p> <ul style="list-style-type: none"> <li>▪ Not hermetically sealed</li> <li>▪ Gets oxidised resulting in frequent failure of heaters</li> </ul> <p>Heater grill</p> <ul style="list-style-type: none"> <li>▪ Fixed grill</li> <li>▪ Reflector cannot be cleaned</li> <li>▪ Increases reflector blackening</li> <li>▪ Unstable heater output</li> </ul> <p>230 volts AC</p> <ul style="list-style-type: none"> <li>▪ Exposed to patient directly</li> <li>▪ Unsafe, threat to patient from sparks</li> </ul>	
2.	<p>Dental Equipment Dental chair with Drill and Accessories Mode Type. AL-398A Computer controlled Integral Dental Unit -4 whole high speed air turbine hand piece -4 whole low speed air motor hand piece -luxurious tempered glass spittoon - Auto spittoon flush and cup filler control system 24V Noiseless DC Motor Chair -3-way syringe(cold/Hot water) -operating light -Saliva ejector with high and low suction - instrument tray with air brake - Water tank( air pressurized) built out - 338 A operating stool -Endoscope system - ultrasound scalar -curing light -medical compressor</p>	10
3.	<p>Shadow less mobile lamp Size 20"-4 reflector Optional-plain/halogen OR -Single dome-300mm dia. -Adjustable Height-1470mm-2000mm - Movable mounted on4 ball wheels - halogen lamp-12V A.C.50 W single phase - light output-45000 Lux</p>	10
4.	<p>Labour Examination table with stirrup - (MI 44) - L1820mm x W 560mm x H 815mm</p>	6

	<ul style="list-style-type: none"> <li>- CRCA top</li> <li>- U notch with foot end folding and detachable</li> <li>- Lithotomy bar and step stool</li> <li>- Pre treated &amp; epoxy powder coated</li> </ul> <p>OR</p> <p>(MI26)</p> <ul style="list-style-type: none"> <li>- Labour table with foot end folding</li> <li>- Size- L1830mm x W870mm x H 88mm</li> <li>- Fram50mmx25mm rectangular stainless steel tube</li> <li>- Tow section stainless steel with head low/up position by screw mechanism</li> <li>- Detachable &amp; folding foot end</li> <li>- U notch with drainage tray, available in full stainless steel top &amp; epoxy powder coated</li> </ul>	
5.	<p>Newborn resuscitation kit</p> <ul style="list-style-type: none"> <li>- 50ml silicon bellow</li> <li>- Non re breathing valve with 40cm/H2O pressure release</li> <li>- Child mask</li> <li>- Patient connector 15mm in size/22mm outside diameter</li> <li>- Oxygen reservoir bag with valve</li> <li>- 1 metre oxygen tubing</li> </ul> <p>And</p> <ul style="list-style-type: none"> <li>- Adult resuscitation kit</li> <li>- 1000 ml silicon bellow</li> <li>- Non re breathing valve</li> <li>- Adult mask</li> <li>- Patient connector 15mm in size/22mm outside diameter</li> <li>- Oxygen reservoir bag with valve</li> <li>- 1 metre oxygen tubing</li> </ul>	1
6.	Vira Tech UV-400(sterilizer of room)	1
7.	Dental chair ( included in common list)	1
8.	Dental X-Ray Dental X-Ray (OPG model) Manual & Digital	1
9.	Semi auto analyzer	1
10.	<p>Phototherapy machine (Chamber ECO) UVA 12-UVB 12) Whole Body Phototherapy machine (12 UVA+12 NB UVB)</p> <ul style="list-style-type: none"> <li>▪ Less space requirement</li> <li>▪ Selector switch for UVA/NBUVB panels</li> <li>▪ Tubes are imported from Phillips Holland</li> <li>▪ Unit Cabinet made of high quality MS with power coated</li> <li>▪ Microprocessor base digital timer for accuracy</li> <li>▪ Independent hour metre for each panel</li> <li>▪ Individual MCB switch on each panel</li> <li>▪ Important mirror type reflectors ensure maximum irradiation</li> <li>▪ Spike guard protection</li> <li>▪ MCB9 (Main Circuit Breaker) ensure max safety for the patient</li> <li>▪ ELCB (Early Leakage Circuit Breaker) for patient safety</li> <li>▪ UV protected glass can be provided for patient and operator</li> </ul>	

	<ul style="list-style-type: none"> <li>▪ Fully insulated circuit base</li> <li>▪ Cooling fans provided</li> <li>▪ Open top unit enables maximum ventilation</li> <li>▪ High quality UV chokes for maximum life for tubes</li> <li>▪ Separate control unit for easy operation</li> </ul> <p>Technical data</p> <ul style="list-style-type: none"> <li>▪ Types of lamp : 12 UVA + 12 NVUVB</li> <li>▪ No. of lamp : 24</li> <li>▪ Lamp Wattage : 100 watt each</li> <li>▪ Electronic Time Selector : Micro control based</li> <li>▪ Reflector : Aluminium Mirror type</li> <li>▪ Operating Voltage : 220/240 Volt 50 Hz 15 AMP</li> <li>▪ Power Consumption : 6KVA</li> <li>▪ Stabilizer required : 6KVA Sevo Voltage Stabilizer</li> </ul>	
11.	Infant Radiant warmer ( in common List)	1
12.	<p>Pulse Oxymeter (Fingertip Pulse Oxymeter POX 100)</p> <p>Key features:</p> <ul style="list-style-type: none"> <li>▪ Simple and convenient to use</li> <li>▪ Low Power Consumption</li> <li>▪ Equipped with functional switch</li> <li>▪ Automatic Power off when no signal is detected for 8 seconds</li> <li>▪ Visual indication for Low Battery</li> </ul> <p>Technical Specifications:</p> <ul style="list-style-type: none"> <li>▪ Display Format: SpO% 35-99% <ul style="list-style-type: none"> <li>• Pulse Ratio 30-250 BPM</li> <li>• Pulse Wave LCD bar graph display</li> <li>• Battery Voltage Low Indication LCD Signals</li> </ul> </li> <li>▪ Resolution: +/- % for SpO% and +/-1 BPM for Pulse Ratio</li> <li>▪ Measurement Accuracy: +/- 2% in stage of 70%-99%</li> <li>▪ Power Consumption : less than 20mA</li> <li>▪ Power Requirement: 2AAA Size 1.5V Alkali Battery</li> <li>▪ Measurement Performance in weak Feeling Condition: Oxyhaemoglobin and Pulse Value can be shown correctly when pulse feeling ration is 0.4%</li> <li>▪ Interference Resistance Capacity against Ambient Light: Deviation is smaller than +/- 1%</li> <li>▪ Display Format: LCD</li> <li>▪ Weight: 70g (With Batteries)</li> <li>▪ Size: 67 mm (L) x 30 mm (H) x 33 mm (W)</li> </ul> <p>Other Models:</p> <ol style="list-style-type: none"> <li>1. Fingertip Pulse Oxymeter POX 100L</li> <li>2. Fingertip Pulse Oxymeter POX 100C</li> </ol>	1
13.	<p>X-ray Machine digital</p> <ol style="list-style-type: none"> <li>1. 300 mA</li> <li>2. 500 mA + IITV</li> <li>3. 100 mA Portable</li> <li>4. Digital Radiography</li> </ol> <ul style="list-style-type: none"> <li>▪ Kodak film</li> <li>▪ Konica film</li> </ul>	1 each

14.	OT focus light (included in common list)	2
15.	<p>Suction machine (Suction Apparatus S-VAC 2000) Electric</p> <p><b>Sailent features</b></p> <ul style="list-style-type: none"> <li>▪ Compact design.</li> <li>▪ Low noise level</li> <li>▪ Maintenance Free</li> <li>▪ Easy to handle</li> <li>▪ Continuous duty</li> <li>▪ Trolley Dry compressed Air</li> <li>▪ Low running cost</li> <li>▪ Capacity 45 ltrs./min air displacement (double Head)</li> </ul> <p>Specifications:</p> <ul style="list-style-type: none"> <li>▪ Trolley – Powder coated MS</li> <li>▪ Standard motor – ¼ HP Single phase</li> <li>▪ Supply voltage – 220V AC</li> <li>▪ Rated current – 2.5 Amp.</li> <li>▪ Creating vacuum – 600 mm to 700 mm of Hg</li> <li>▪ Cold sterilizable glass bottle 2 Nos. of 1600 cc</li> <li>▪ Noiseless function</li> </ul> <p>Manual (foot suction)</p> <ul style="list-style-type: none"> <li>▪ Compact, light weight, easy to handle and operate</li> <li>▪ Durable rubber bellow</li> <li>▪ Long lasting stainless spring to provide friction free pumping</li> <li>▪ Completely autoclavable polycarbonate vacuum jar with lid (500 ml capacity)</li> <li>▪ Scratch proof powder coated frame</li> </ul>	<p>2</p> <p>4</p>
16.	<p>Stethoscope (littmann)</p> <p>Paediatric</p> <p>Adult</p>	<p>1</p> <p>4</p>
17.	<p>Otoscope</p> <p>Otoscope plastic body pouch pack</p> <p>Plastic body mini fibre optic otoscope along with 7 pieces of additional reusable specula of 2 x 4 mm</p>	2
18.	Ophthalmoscope	2
19.	<p>ECG machine</p> <ul style="list-style-type: none"> <li>- 3.8" large LCD display 12 lead ECG machine</li> <li>- Wave form synchronously</li> <li>- Input circuit : floating: protection against defibrillation effect</li> <li>- Lead : standard 12 leads</li> <li>- Patient current leakage : &lt;10A or&gt;50MQ</li> <li>- Input impedance : &gt;50W</li> <li>- Calibrating voltage : 1mV+-3%</li> <li>- Digital sampling : 12 bits( simultaneously acquisition of 12 lead</li> </ul> <p style="text-align: center;">ECG data</p> <ul style="list-style-type: none"> <li>- Frequency response : 0.05Hz~150Hz or 165(-3dB)</li> <li>- Time constant : &gt;3.2 seconds</li> <li>- CMMR : &gt;100dB( with filter)</li> </ul>	1

	<ul style="list-style-type: none"> <li>- EMG FILTER : 40Hz(-3dB)</li> <li>- DFT FILTER : Automatic adjustment</li> <li>- Sensitivity : 2.5, 5, 10, 20mm/mV+-3%</li> <li>- Paper speed : 6.25, 12.5, 25, 50mm/s, +-3%</li> <li>- Recording system : thermal –array</li> <li>- Recording paper : 112mm,80mm,20mm high speed rolling</li> <li>- Input circuit current : &lt;50nA</li> <li>- Power supply : AC220V/110V, 50/60HZor DC 20V, 2A /14.4V 2Ah in built</li> </ul> <p style="text-align: center;">Rechargeable L ion battery</p> <ul style="list-style-type: none"> <li>- Depolarization voltage : +- 400mV or +-450 mV</li> <li>- Noise level : &lt;15Uvp-p</li> <li>- Safety level IEC class : 1,type CF</li> <li>- Dimension : 300x 230x 65mm or285x250x63mm</li> <li>- Dimension carton : 380mmx520mm 195mm</li> <li>- Net wt. : 2.75 Kg or 1.7 Kg</li> <li>- Gross wt. : 4.5 Kg or 4.2 Kg</li> </ul>	
20.	<p>Autoclave HP vertical (2 Bin) Vertical Sterilizer (GMP Model) – double walled Technical specification: The Autoclave shall be ISI marked IS:3829-1978 with latest amendments. The inner and outer chamber is made of high-grade stainless steel sheet (SS-304); the strong and sturdy lid is also made of high grade stainless steel plate. (SS-304). The unit is fitted with very efficient radial locking system made of stainless steel bars (SS-304), steel, chrome plated, with paddle lifting device. Supplied complete with stainless steel basket. The unit is fitted with safety valve, pressure release valve, water level indicator, cord and plug. To operate on 220/230 volts, 50 Hz AC supply. Documentation: Supplied complete with operation protocol, installation protocol and design qualification certificates along with the necessary hydro test certificates and other calibration certificates Mode of Heating: Electrically heated by special flange type/immersion type heaters bearing ISI mark, to operate on 220/230 volts, 50Hz AC supply</p>	2
21.	<p>OT Table S-Con Hydraulic Operation Theatre Table <b>Features</b></p> <ul style="list-style-type: none"> <li>• Heavy Duty Hydraulic Operating Table having side control</li> <li>• Stainless Steel Top Five Sections</li> <li>• Jerk Free Smooth Operating Facilities</li> <li>• Table Base Covered with Stainless Steel Sheet</li> <li>• Appropriate for Gynaecology, Urology, Endoscopy, Laparoscopy &amp; General Surgery</li> <li>• Detachable Head &amp; Leg Section</li> <li>• Easily Moveable &amp; Floor Locking System</li> </ul>	1

	<ul style="list-style-type: none"> <li>• Table Top-turnable to sitting position</li> </ul> <p><b>Parts:</b></p> <ol style="list-style-type: none"> <li>1. Base</li> <li>2. Arm Rest</li> <li>3. Head Section</li> <li>4. Anaesthetic Screen</li> <li>5. Shoulder Supports</li> <li>6. Back Section</li> <li>7. Chest &amp; Waist Supports</li> <li>8. Seat Section</li> <li>9. Kidney Section</li> <li>10. Leg Section</li> <li>11. Knee Crutches</li> <li>12. Foot Rest</li> <li>13. Foot Paddle</li> <li>14. Sponge Mattress</li> </ol> <p><b>Operating Figures</b></p> <p>Specifications</p> <ul style="list-style-type: none"> <li>▪ Table Top – (180-195) x (50-55) cm</li> <li>▪ Height – Min 75 cm &amp; Max 90 cm</li> <li>▪ Hydraulic Lift – 15 cm Min</li> <li>▪ Trendelenburg - <math>\pm</math> (29° to 31°)</li> <li>▪ Tilting – Lateral - <math>\pm</math> (18° to 22°)</li> <li>▪ Head Section 45° up &amp; 90°Dn</li> <li>▪ Leg Section 90°Dn</li>   <li>▪ Back Section Chair Form</li> </ul> <p><b>Standard Accessories</b></p> <ul style="list-style-type: none"> <li>▪ Anaesthetic Screen – 1 pc</li> <li>▪ Foot Rest – 1 pc</li> <li>▪ Knee Crutches – 2 pcs (Optional at extra cost)</li> <li>▪ Sponge Mattress 50 mm Thick with Rexin Cover</li> <li>▪ Chest &amp; Waist Supports – 2 pcs</li> <li>▪ Shoulder Supports – 2 pcs</li> <li>▪ Arm rest – 2 pcs</li> <li>▪ Lithotomy Leg Holder – 2 pcs</li> </ul>	
22.	<p><b>Diathermy machine</b></p> <ul style="list-style-type: none"> <li>- Surgical Diathermy, Microprocessor controlled electrosurgical unit,</li> <li>- With all standard accessories</li> <li>- Monopolar and bipolar modes</li> <li>- Soft, force and spray coagulation.</li> <li>- 4 steps of Haemostasis</li> <li>- Patient plat test function</li> <li>- Bipolar coagulation with auto start.</li> </ul> <p><b>Monoplar cut</b></p> <ul style="list-style-type: none"> <li>- Waveform of HF Voltage : unmodulated sinusoidal</li> <li>- Frequency of HF voltage : 350kHz(+/-20kHz)</li> <li>- Haemostatic(blend) : 4 steps coagulation degree via front panel</li> </ul> <p style="text-align: center;">Keys</p> <ul style="list-style-type: none"> <li>- Maximum HF Voltage : 1.7 KV peak to peak</li> </ul>	4



	<ul style="list-style-type: none"> <li>- HF output for SC : 400 Watts at 200 to 700 ohms</li> <li>- Power settings Keys : 1-400 watts in increment of 1 watt by Up/down soft</li> <li>- Acoustic signal : Continuous tone</li> </ul> <p><b>Monopolar soft coagulation</b></p> <ul style="list-style-type: none"> <li>• Waveform of HF Voltage : Pulse modulated</li> <li>• Frequency of HF Voltage : 5000kHz(+/-25kHz)</li> <li>• Maximum HF Voltage : 3000 V peak to peak</li> <li>• HF output : 120 watts at 300ohms</li> <li>• Power setting keys : 1 to 120 watts in increment of 1 watt by Up/down soft</li> </ul> <p>Monopolar Force coagulation</p> <ul style="list-style-type: none"> <li>• Waveform of HF Voltage : Un modulated Sinusoidal</li> <li>• Frequency of HF Voltage : 350kHz(+/-20kHz)</li> <li>• Maximum HF Voltage : 530 V peak to peak</li> <li>• HF output : 120 watts at 100 to 200ohms</li> <li>• Power setting keys : Modulated tone</li> </ul>	
23.	Shadow less lamp( included in common list)	2
24.	Focus lamp ordinary/Shadow less mobile lamp Single-dome-300mm dia.adjustable height-1470mm-2000mm Movable mounted on 4 ball wheels Halogen lamp-12V AC.50 w single phase	2
25.	Anaesthetic laryngoscope magils with four blades	3
26.	Endo tracheal tube set	2
27.	Magils forceps (two size)	2
28.	Connector set of six for E.T.T.	2
29.	Tubes connecting for E.T.T.	2
30.	N2O cylinder for Boyles	10
31.	Vaginal Examination set	5
32.	Thumb forceps medium size	10
33.	MR syringe	4
34.	Ophthalmoscope(specialist model)	2
35.	MTP set	2
36.	Multi parameter monitor ,6 channel monitor with 10.4" high resolution TFT display suitable for Adult,Padiatric & Neonatal applications, Night mode & user reminder alarm,users selectable Display formats 168 hours(7 days) graphical tubular trends	1
37.	Puva apparatus for Dermatology Derma India whole Body Phototherapy chamber (ECO) SPIEGEL SEIES UVA 12-UVB12,Dimenson outer(1955Hx1070x1070D)	1
38.	T Handle	1
39.	Operation table ordinary, not hydraulic with all position	1

**DECLARATION**  
**On**  
**Manufacturing facilities**

**Tender enquiry No.** \_\_\_\_\_

For supply of \_\_\_\_\_

1. Name of the tenderer :
2. Full Postal Address :
3. Telephone No./Fax No. :
4. Email address :
5. Date of inception of business :
6. Registration no. & Date :
7. Issued by :
8. Valid till :
9. Details of manufacturing activity :  
& item wise capacity
10. Detail of After Sales Service & AMC facilities available locally

Name of the Agency:

Full Postal Address:

Phone / Fax / E-mail:

11. Name of person responsible for 10 above

Sl. no.	Name	Designation	Age	Residential Address
---------	------	-------------	-----	---------------------

12. Name of Govt. Departments/ Pvt. Institutions to which the tenderer already supplied the items with quantity, value and supply period As per enclosure
13. Has the tenderer ever been black listed by any govt. agency? If yes, give details.
14. Are any cases pending in the court related to any supplies? If yes, give details
15. Does the firm have the adequate facilities for inspection and quality control? Please give details

I, \_\_\_\_\_ Prop./partner/Director of M/s

\_\_\_\_\_

Hereby declare that the information given in this form is true and correct to the best of my knowledge & belief.

I/we agree to the tender Inviting Authority forfeiting the Earnest Money Deposit and/or Performance Security Deposit and blacklisting us for a period of 5 years, if any information furnished by us proved to be false at the time of inspection and non – compliance with terms and conditions of the contract

I offer to supply the items mentioned in the schedule (enclosed in price bid) at the rates quoted therein. I agree to hold this offer for one year after finalization of rate contract.

Dated: Signature\_\_\_\_\_

Name of tenderer\_\_\_\_\_

Address\_\_\_\_\_

\_\_\_\_\_

**PROFORMA FOR PERFORMANCE STATEMENT**  
**(FOR A PERIOD OF LAST 3 YEARS)**

Name of firm \_\_\_\_\_

Sl.	Name of the product	Year	Quantity supplied.	Value.	Name and full address of the purchaser
	1	2	3	4	5
1.					
2.					
3.					

Signature and seal of the Tenderer \_\_\_\_\_

## ANNEXURE-III

### ANNUAL TURN OVER STATEMENT

The Annual Turnover of M/s\_\_\_\_\_ for the past three years and concurrent commitment for the current financial year are given below and certified that the statement is true and correct.

---

<i>Sl.No.</i>	<i>Year</i>	<i>Turnover in Lakhs (Rs)</i>
---------------	-------------	-------------------------------

---

1.	2009-10	-
2.	2010-11	-
3.	2011-12	-

---

Total - Rs. \_\_\_\_\_ Lakhs.

---

Average turnover per annum - Rs. \_\_\_\_\_ Lakhs.

#### Concurrent Commitment

Sl. No.	Contract Ref.	Purchaser	Total Contract Value	outstanding Value	Estimated Delay in completion date

**UNDERTAKING**

To

Tender enquiry No. \_\_\_\_\_

For supply of \_\_\_\_\_

Sir,

1. I, Shri \_\_\_\_\_, on behalf of M/s \_\_\_\_\_ having registered office at \_\_\_\_\_, do hereby declare that I have gone through the terms and conditions mentioned for the above and undertake to comply with all tender terms and conditions. The rates quoted by me/us are valid and binding on me/us for acceptance for a period of one year from the date of award of contract to us.
2. I/We undersigned hereby bind myself/ourselves to the Office of .....to supply .....  
The rates quoted by me/us for the items tendered for are specified against each. It is certified that rates quoted are lowest quoted for any institution in India and not higher than the MRP/ prevailing market rate.
3. The articles shall be strictly as per specification and of the best quality as per requirement of the institution. The decision of the Office of ..... (hereinafter called the said Purchaser) as regards to the quality and specification of article shall be final and binding on me/us.
4. We undertake and confirm that

- h) COMPREHENSIVE WARRANTY: The medical equipment, supplied shall be covered under comprehensive warranty for a period of 3 years from the date of commissioning and any maintenance or repair arising during this period of 3 years including replacement of parts if any, shall be carried out by us at our own cost without any cost liability on the purchaser.
- i) ANNUAL MAINTANANCE CONTRACT (AMC): We have quoted the %age rates for the Comprehensive Annual Maintenance of the equipment tendered, after expiry of warranty period of 3 years, for a period of 5 years, in the price bid. We note that **Price bids without AMC rates shall not be considered for evaluation.**

We have also quoted annual maintenance charges for next 5 years both on comprehensive basis including spares and only for services without spares separately. These rates shall be valid for acceptance by the purchaser, till 5 years from commissioning of the equipment.

- j) Authorization from Equipment Manufacturer: We shall furnish authorization from the manufacturer legally enforceable undertaking to the Purchaser in appropriate format assuring full guarantee/ warranty obligations valid for a period of 3 years from the date of commissioning of the supplied equipment. We also confirm that we shall provide Comprehensive Annual Maintenance Services directly or through the manufacturer for at least 5 years after expiry of warranty period of 3 years, in terms of Annual Maintenance Contract to be entered into between us/ Manufacturer and the Purchaser. Further we agree to provide Annual Maintenance services for next 5 years by similar arrangement and in line with (b) above.

- k) Response Time for Breakdown Calls: Maintenance personnel shall attend to the breakdown within 72 hrs of reporting the fault.
  - l) Preventive Maintenance: During the warranty period and AMC, in addition to attending break down calls, there shall be regular quarterly preventive maintenance visits.
  - m) Operation & Maintenance Manuals: We shall furnish for each unit 2 sets of operation & maintenance manuals along with circuit diagram and spare parts list of the equipment.
  - n) Installation, Commissioning & Warranty Service: Supply, installation, commissioning and successful trial run of the equipment & subsequent, warranty servicing shall be done at site by us or our authorized agent at no extra charge.
  - o) Training: If required, free demonstration, operation and maintenance training of the assigned personnel at site shall be provided by us.
5. We agree to the conditions of the tender under which the EARNEST MONEY DEPOSIT and PERFORMANCE SECURITY DEPOSIT shall be forfeited by us.
  6. We hereby undertake to pay the **penalty** as per the terms and conditions of the contract for delayed supply of the ordered items.
  7. We agree to accept the amount of the bill to be paid by the purchaser after completion of all codal formalities and should any amount of the bill found by the purchaser/auditors to have been over-paid; the amount so found shall be refunded by me/us.
  8. We hereby undertake to supply the items during the validity of the tender as per direction given in supply order within the stipulated period.
  9. The tender inviting authority has the right to accept or reject any or all the tenders without assigning any reason.



10. We understand all the terms and conditions of the contract and bind myself/ourselves to abide by them.

11. We hereby declare that there is no vigilance/CBI or court case pending/contemplated against us at the moment.

SIGNATURE :

NAME & DESIGNATION :

DATE :

NAME & ADDRESS OF  
THE FIRM :

SEAL :

**UNDERTAKING ON FRAUD AND CORRUPTION**

We M/s ..... Do hereby undertake that, in competing for (and, if the award is made to us, in executing) the subject contract for supply of ..... under tender reference no. .... Dt ..... We shall strictly observe the laws against fraud and corruption in force in the country.

Sd/-

Signature of proprietor/Partner/Director  
Designation:

Seal:

**MANUFACTURER'S AUTHORIZATION FORM**

**No.** \_\_\_\_\_ **Dated** \_\_\_\_\_

**To,**

Dear Dir,

Bid Ref. No. \_\_\_\_\_

We \_\_\_\_\_ who are established and reputable manufacturers of \_\_\_\_\_ having factories at \_\_\_\_\_ Registered office at \_\_\_\_\_ possessing Manufacturing Licence No. \_\_\_\_\_ dated \_\_\_\_\_, Valid Upto \_\_\_\_\_, hereby authorize M/s. \_\_\_\_\_

\_\_\_\_\_ (Name and address of Representative), to submit a bid and subsequently negotiate and sign the contract with you against the above mentioned tender for the following items quoted.

- 1.
- 2.
- 3.

(Attach separate sheet if necessary)

We hereby extend our comprehensive Guarantee/ warranty for a period of 3 years from the date of commissioning of the equipment and commit to provide annual maintenance services for a further period of minimum 5 years as per the tender conditions for the items offered for supply against the tender submitted by the above

firm. After expiry of this period, Annual Maintenance Services for another 5 years either on comprehensive basis (including spares) or for services only shall also be provided, if the purchaser so desire.

Yours faithfully,

(Name)

for and on behalf of M/s. \_\_\_\_\_

(Name of Manufacturer)

Note: This letter of authority should be on the letterhead of the manufacturing concern and should be signed by a person competent and having the power of attorney to bind the manufacturer.

**ANNEXURE-VII**

**FOR COVER 'A' –UNPRICED BID**

**Tender No.**\_\_\_\_\_

**A. List of Equipments / Items**

Schedule No.	Sr. No.	Description	Qty	Price Quoted / Not Quoted	Technical Specification attached / Not attached
--------------	---------	-------------	-----	---------------------------	---

**NAME OF THE PURCHASER****AGREED TERMS & CONDITIONS**

Tender No. &amp; Date \_\_\_\_\_

**A. Details of Bidder**

Bidder

Offer Ref

Telephone No.

Fax No.

Contact Person:

Signature:

E-mail:

**B. Definitions**

1. "Purchaser" means ..... or his authorized representative.
2. "Bidder" or Tenderer means a person or firm or company who has made an offer for supply of goods and /or service as per tender.
3. "Vendor" or "Supplier" means a person or firm or company, to whom the order is addressed for supply of goods and /or services.
4. "Site" means the premises of the purchaser or any other place as decided by the purchaser.

**NOTE: The questionnaire below must be duly filled in and should be enclosed with un-priced bid, Bid 'A'. Clauses confirmed here under should not be repeated. All commercial terms and conditions should be indicated in this format. If necessary, details including deviations to the terms and conditions of the bid document, if any, should be enclosed as annexures to this questionnaire.**

<b>Sl. No.</b>	<b>Description</b>	<b><u>Vendor's Confirmation</u> (Confirmed/Noted/Deviation/ furnished separately)</b>
	<b>C. Technical</b>	
1.	Confirm that you meet the eligibility criteria as per bid document and have furnished relevant documents.	
2.	Confirm acceptance of Technical Specification and scope of supply as per Tender Document.	
3.	In case of deviations, confirm that the same have been highlighted separately.	
4.	Confirm that equipment catalogue and technical data, wherever applicable, have been enclosed.	
5.	Confirm that all certificates/ deviations furnished.	
6.	Confirm that Earnest Money Deposited (EMD) as per bid document, for each schedule quoted, enclosed.	
	<b>D. Commercial</b>	

1.	It is noted that any deviations to the commercial terms and conditions shall lead to loading of prices or rejection of offer.	
2.	Confirm that the quoted prices are on duly installed and commissioned at site basis including packing & forwarding (P & F) all duties and taxes viz. Excise Duty, Sales Tax/VAT, freight, insurance, installation & commissioning at site.	
3.	Confirm furnishing of detail price break-up of each item showing all components of cost separately with basic price of equipment and other costs on %age of basic price to arrive at landed price in D2 above.	
4.	It is noted that the statutory variations in taxes and duties within the Contractual delivery period shall be borne by the purchaser.	
5.	If there is any variation or fresh imposition of Excise Duty at the time of supply due to various reasons, including turn-over, confirm that the same shall be borne by you.	
6.	It 5 above is not acceptable, advise maximum possible rate of ED changeable, which shall be loaded to you price.	
7.	Confirm that in case of delay on you account any new or additional duties and taxes imposed after the contractual delivery date shall be to you account. This will be in addition to Price Reduction for Delay in Delivery.	
8.	Confirm acceptance of Price Reduction Schedule for delay in Delivery @ 0.5% of delayed value of goods per week of delay or part thereof subject to maximum of 10% of the total order of value.	
9.	Indicate Firm Delivery Period, which shall be counted from the date of placement of order and date of commissioning and acceptance of the installed equipment by purchaser's representative shall be considered as the delivery completion date.	
10.	Confirm acceptance of relevant payment terms specified in the bid document.	
11.	Force Majuere—Delivery period, price reduction termination etc are subject Force Majuere Condition as stipulated in the bid document.	
12.	Confirm that the quoted prices shall remain firm & fixed till complete execution of the order.	
13.	a) In case you are a manufacturer confirm that the prices quoted are not higher in any respect than quoted / agreed prices to any Govt. or Public Medical Institution. b) In case you are a dealer/ distributor / authorized agent, confirm that the prices quoted are as per manufacturer's price list with appropriated discount (Copy of Price list to be enclosed).	
14.	a) Confirm that all inspection & testing charges	

	<p>including 3<sup>rd</sup> party inspection (if required) included in the price.</p> <p>b) Quality Control Reports and Test Certificates, whenever applicable, shall be handed over to the purchaser along with the equipment.</p>	
15.	<p>a) Confirm that erection, commissioning, trial run and handing over to the purchaser, after successful commissioning is your responsibility at no extra cost.</p> <p>b) Specify facilities (if any), like water supply/ electric power to be provided by the purchaser for commissioning of the equipment.</p>	
16.	<p>Packing / forwarding, transportation, loading/ unloading and insurance are your responsibility.</p> <p>However, to protect the equipment from physical damages and/or deterioration due to weather during transit, you are to ensure proper packing &amp; lifting arrangement. Pl confirms compliance.</p>	
17.	<p>The material / equipment shall be guaranteed against any and all defects in design, workmanship, material &amp; performance for a period of 3 years from the date of commissioning and handing over to the purchaser. Should any defect detected or develop during the guarantee period, it shall be remedied promptly free of cost by the supplier and all expenses including transportation of goods necessitated for such repair and replacement shall be done by the supplier. Pl confirm acceptance.</p>	
18.	<p>Confirm that Performance Bank Guarantee shall be furnished for 10% of the total order value valid till guarantee period of the goods.</p>	
19.	<p>Confirm your acceptance to enter into an agreement with the purchaser, for Comprehensive Annual Maintenance Contract for at least 5 years after expiry of guarantee / warranty period of the goods.</p> <p>Confirm that the AMC changes on %age basis have been quoted in you price bid.</p>	
20.	<p>Confirm acceptance of Part order.</p>	
21.	<p>Confirm acceptance of Repeat order within 12 months from the date of basic order at same price and terms &amp; conditions.</p>	
22.	<p>In case of material having shelf life, confirm that you have declared the same with the expiry date. Also confirm that such materials shall be dispatched within 30 days from the date manufacture.</p>	
23.	<p>It is noted that the purchaser would disown any responsibility / liability toward irregularity, contravention or infringement of any statutory regulations including those of patent, on manufacture or supply of goods covered by the order.</p>	



24.	Terms & Conditions indicated in this format shall not be repeated in the bid. Terms & Condition indicated elsewhere and contradicting those in this format shall be ignored. Confirm compliance.	
25.	Confirm that you shall observe the highest standard of ethics during bidding and in case favoured with an order, during execution of the order, without resorting to any fraud, corruption and/or coercion.	
26.	Confirm that the offer shall be valid for a period of 90 days from the date of bid opening.	

**ANNEXURE-X**

**FOR COVER 'B' – PRICED BID**

**Tender No.**\_\_\_\_\_

**A. List of Equipments/ Items**

**SCHEDULE OF RATES**

Schedule No.	Sr. No.	Description	Quantity	Total landed cost including installation & warranty maintenance
--------------	---------	-------------	----------	---

**FORM OF CONTRACT AGREEMENT**

THIS CONTRACT AGREEMENT is made

The ..... day of ....., year.....

BETWEEN

(1) Name and Address of the Purchaser:

(2) Name and Address of the supplier:

WHEREAS the Purchaser invited bids for certain goods and ancillary services, viz., [insert: brief description of goods and services] and has accepted a bid by the Supplier for the supply of those goods and services in the sum of [insert: contract price in words and figures] (hereinafter called “the Contract Price”)

NOW THIS AGREEMENT WITNESSES AS FOLLOWS:

1. In this Agreement words and expressions shall have the same meaning as are respectively assigned to them in the Condition of Contract referred to.
2. The following documents shall constitute the Contract between the Purchaser and the Supplier, and each shall be read and constructed as an integral part of the Contract:
  - (a) This Contract Agreement
  - (b) Special Condition of Contract
  - (c) General Condition of Contract
  - (d) Technical Requirements ( including Technical Specifications)
  - (e) The Supplier’s bid and original Price Schedules

(f) The Purchaser's Notification of Award

(g) [Add here: **any other documents**]

3. In consideration of the payments to be made by the Purchaser to the Supplier as hereinafter mentioned, the Supplier hereby covenants with the Purchaser to provide the Goods and Services and to remedy defects therein in conformity in all respects with the provisions of the Contract.
4. The Purchaser hereby covenants to pay the Supplier in consideration of the provision of the Goods and Services and the remedying of defects therein, the Contract price or such other sum as may become payable under the provisions of the Contract at the times and in the manner prescribed by the Contract.

For and on behalf of the Purchaser

Signed: \_\_\_\_\_

In the capacity of [insert: **title or other appropriate designation**]

In the presence of \_\_\_\_\_

For and on behalf of the supplier

Signed: \_\_\_\_\_

In the capacity of [insert: **title or other appropriate designation**]

In the presence of \_\_\_\_\_